



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 30, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 6 ADDRESS: SR97 MP265

LOCATION Brewster 91

COUNTY: Okanogan

MAKE: RiceLake

TYPE Trk Scale

ACCURACY CLASS: IIIIL

MODEL: IQ310

ID. NO: 91441332

CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Digital

PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20,000	20,000	0				
CENTER	20,000	20,000	0				
SEC 2	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer

Date: July 1, 2010

Print name: David Cromer

State of Washington

County of Shenston

Signed and sworn to (or affirmed before
me on July 1, 2010 by

Brenda A. Biscay

(Signature of notary public)

Brenda A. Biscay

Printed Name of Notary

My appointment expires 3.12.12

